

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUN 22 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130121</u> ✓	
Name <u>John Pernell</u>		ID# <u>6260</u>	Date <u>17 Jun 09</u> ✓
A	Agency <u>Fairbanks Airport Police</u>		Phone # <u>474-2555</u>
Instrument Location <u>Airport Police Station, Fairbanks, AK</u> <u>5195 Brumbaugh Ave. Fairbanks, AK 99709</u>			
B	Alco S/N <u>57447</u> ✓	Target Value <u>.079</u>	High Pressure <u>730 psi.</u>
Alco Test Values		<u>.079</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>John Pernell</u> ✓			
(OVER)			

JMB 6/25/09  
6/25/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of June, 2009.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

JUNE 17, 2009

OPERATOR'S NAME:

PERNELL/JOHN/T

OPERATOR'S NUMBER: 6260

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: FBK6

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: 57447

### BREATH ANALYSIS

.079 ADJUSTED FOR 29.55 in		
ALCO TARGET	.078	02:19
BLANK TEST	.000	02:20
INTERNAL STANDARD	VERIFIED	02:20
ALCO TV 29.55 in	.079	02:21
BLANK TEST	.000	02:21
SUBJECT SAMPLE	.000	02:22
BLANK TEST	.000	02:23
ALCO TV 29.55 in	.079	02:23
BLANK TEST	.000	02:24

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

JUNE 17, 2009

TIME 02:26

### DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	46c
BAROMETER:	29.57 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢